

# G. LUFFT Mess- und Regeltechnik GmbH

akkreditiert durch die / accredited by the

## Deutsche Akkreditierungsstelle GmbH

als Kalibrierlaboratorium im / as calibration laboratory in the

## Deutschen Kalibrierdienst



Kalibrierschein  
Calibration certificate



Deutsche  
Akkreditierungsstelle  
D-K-15202-01-00

Kalibrierzeichen  
Calibration mark

9476
D-K- 15202-01-00
2018-02

Gegenstand Object	<b>temperature transmitter</b>
Hersteller Manufacturer	<b>NRG Systems</b>
Typ Type	<b>110 S</b>
Fabrikat/Serien-Nr. Serial number	---
Auftraggeber Customer	<b>HydroWind BVBA Veldkantstraat 119 B-1850 Grimbergen (Brussels)</b>
Auftragsnummer Order No.	<b>REP 6526A</b>
Anzahl der Seiten des Kalibrierscheines Number of pages of the certificate	<b>3</b>
Datum der Kalibrierung Date of calibration	<b>15.02.2018</b>



Dieser Kalibrierschein dokumentiert die Rückführung auf nationale Normale zur Darstellung der Einheiten in Übereinstimmung mit dem Internationalen Einheitensystem (SI). Die DAkkS ist Unterzeichner des multilateralen Übereinkommens der European co-operation for Accreditation (EA) und der International Laboratory Accreditation Cooperation (ILAC) zur gegenseitigen Anerkennung der Kalibrierscheine. Für die Einhaltung einer angemessenen Frist zur Wiederholung der Kalibrierung ist der Benutzer verantwortlich.

*This calibration certificate documents the traceability to national standards, which realize the units of measurement according to the International System of Units (SI).*

*The DAkkS is signatory to the multilateral agreements of the European co-operation for Accreditation (EA) and of the International Laboratory Accreditation Cooperation (ILAC) for the mutual recognition of calibration certificates. The user is obliged to have the object recalibrated at appropriate intervals.*

Dieser Kalibrierschein darf nur vollständig und unverändert weiterverbreitet werden. Auszüge oder Änderungen bedürfen der Genehmigung sowohl der Deutschen Akkreditierungsstelle GmbH als auch des ausstellenden Kalibrierlaboratoriums. Kalibrierscheine ohne Unterschrift haben keine Gültigkeit.

*This calibration certificate may not be reproduced other than in full except with the permission of both the Deutsche Akkreditierungsstelle GmbH and the issuing laboratory. Calibration certificates without signature are not valid.*

Datum Date	Leiter des Kalibrierlaboratoriums Head of the calibration laboratory	Bearbeiter Person in charge
15.02.2018	 Helmut Hager	 Frank Bidmann

Dieser Kalibrierschein ist elektronisch signiert und liegt als Original als PDF-Datei vor.  
*This calibration certificate is electronic signed and exists as original as PDF-file.*

G. LUFFT Mess- und Regeltechnik GmbH  
Gutenbergstraße 20  
DE-70736 Fellbach  
Germany

Tel: ++49(0)711/51822-0  
Fax: ++49(0)711/51822-41  
E-Mail: [info@lufft.de](mailto:info@lufft.de)  
[www.lufft.de](http://www.lufft.de)

Geschäftsführer/Managing directors:  
Dr. Anton Felder  
Jörg Mayer  
Dr. Martin Nicklas  
Axel Schmitz-Hübsch

Amtsgericht Stuttgart  
HRB 721373  
Ust.ID: DE 250580689  
Steuernummer 90490/28336

9476
D-K- 15202-01-00
2018-02

### calibration item

The calibration object is a temperature transmitter.

The calibration object is unscathed.

#### **absolute pressure**

measuring range -40 °C ... 52.5 °C

accuracy +/- 1,24 K

transfer function temperature = (Voltage x 55,55) – 86,38 °C

**Correlation**  
**0,999999952**

**Slope[C/V]**  
**55,708**

**Offset[C]**  
**-86,994**

### reference standard

#### **Temperature**

reference standards PT100 resistance thermometer

reference numbers 006197, 006400, 006401, 006404, 006405

calibration marks 7764-, 7770-, 7771-, 7772-, 7773-D-K-15202-01-00 2017-06

uncertainty of measurement 5 mK...15 mK

#### **Temperature**

reference standard precision temperature measuring instrument

reference number 801062

calibration mark 01-1090-D-K-15186-01-00 2016-10

uncertainty of measurement 3,3 mK...8,3 mK

#### **Other measuring instruments**

reference standard digital multimeter

reference number 801058

calibration mark 2122-D-K-15042-01-00-2017-08

uncertainty of measurement 0,00015 % ... 0,028 % of value

### calibration procedure

#### *Temperature:*

The temperature calibration was accomplished after the DAkkS-DKD guideline "Calibration of resistance thermometers" DAkkS-DKD-R 5-1 from December 2010.

The temperature values ( $t_{90}$ ) refer to the International Temperature Scale of 1990 (ITS-90).

### measurement conditions

#### *temperature:*

climate chamber, medium: air

#### *description:*

The calibration unit was placed in the center of the climate chamber and was completely exposed to the climate conditions. The output signal of the calibration unit was recorded with a multimeter and the software „BenchLink Datalogger 3, Ver. 3.10.00“. The measuring interval was 10 seconds and over 10 minutes the arithmetic mean value was build. The supply voltage was 5 VDC.

#### ambient conditions

temperature in °C: 21,0 ± 1 K  
rel. humidity in %: 30 ± 10 %  
air pressure in mbar: 979 ± 10 mbar

#### calibration results

##### Temperature calibration

Reference standard	Calibration unit			
temperature $t_{90}$ in °C	output signal U in V	temperature calculated $t_{90}$ in °C	measurement deviation $\Delta T_{90}$ in K	uncertainty of measurement U in K
-0,03	1,561	0,33	+0,36	0,21
20,07	1,922	20,39	+0,32	0,15
40,08	2,281	40,33	+0,25	0,13

#### measurement uncertainty

The uncertainty stated is the expanded uncertainty  $U$  obtained by multiplying the standard uncertainty by the coverage factor  $k = 2$ . It has been determined in accordance with DAkkS-DKD-3. The value of the measurand lies within the assigned range of values with a probability of 95%.